



VARTECTOR

Varnish Potential Evaluation System
By MPC (Membrane Patch Colorimetry)

Lubrication Plus 

www.lubricationplus.net
lubricationplus@gmail.com



VARTECTOR

VARTECTOR detects the risk of turbine oil varnish ASTM D7843 (Standard Test Method for Measurement of Lubricant Generated Insoluble Color Bodies in In-Service Turbine Oils using Membrane Patch Colorimetry). The CIE delta E value, the optical colorimetric notation, is used to diagnose the risk of varnish formation. Potential hazards of turbine oil varnish warns that problems such as turbine bearing vibration and temperature rise, filter clogging, poor temperature control, turbine oil degradation may occur.

VARTECTOR diagnoses the potential hazards of varnish and soot in the control oil (phosphate ester Inflammable hydraulic oil) of the EHC system through CIE delta L value and a and b values as optical colorimetric notation, It is possible to diagnose whether the composition of the thermal load is created soot or varnish or whether both carbon and varnish are present and its seriousness. When CIE delta L value is high, it means generation of soot. It means that there is a bubble problem in the system and pump cavitation problem. When a and b values are high, it means a risk of varnish formation. Warning of danger.

Key Features:

- Fully Automation: Unlike existing MPC tester, the process of loading, validating, and testing the samples is done automatically within 10 seconds after patching.
- Display: Patch image, MPC delta E, delta L, a, b value, Trending per machine to be managed.
- Complied to ASTM D7843: MPC tester for laboratory analyzer that meets ASTM D7843 for the first time in the world
- Automatic Diagnostic Reporting: Automatic diagnosis of the measured result (normal a, normal b, caution, warning)
- Automatic Validation: Unlike existing MPCs for existing portable devices, it is built in the device and automatically performs device validation after power on.
- Automatic Self Diagnostic: Perform self-diagnosis of major components such as spectrophotometer and board after power on
- Trending: It is possible to manage the tendency when testing after registering facility information for each machine to be managed in the connector.
- Advanced Software: Perform manual validation, set reference value for management target, trending, save more than 10,000 tables, perform instrument calibration

	Mineral oil - turbine oil	Phosphate ester - Control oil
Formulation	Base Oils (mineral based AP1 group II or III) + additives (antioxidants, rust inhibitors, anti-emulsifiers, anti-foam agents)	Base Oils (phosphate ester synthetic oil) No or little additives
Cause of main degradation	Oxidation	Hydrolysis + oxidation + micro dieseling
Degradation factors	Temperature, deterioration, air, metal particles, water, etc.	Water, oxygen, air bubbles
The contamination by oil degradation (sludge)	Soluble organic acids (i.e., oxides)	Inorganic acid (phosphoric acid), soot (carbide), GEL (gel, generated from acid control filter)
Degraded oil Patch		

Mineral Turbine Oils Application

As the varnish value increases, bearing vibration and temperature hunting occur, and a varnish layer is formed on the surface of the bearing to interfere with the flow of turbine oil, making the cooling action difficult, the early clogging of the filter and the varnish material attached to the cooler causing the cooling operation of the turbine oil to become ineffective and causes many problems

“Oxidation of turbine oil changes fluid properties, and Reduces machine life.”




Normal A	Normal B	Caution	Warning
<15	15 - 29	30 - 40	>40




Progression of oil degradation = Increase of MPC value



Vibration and temperature hunting of bearings



Early clogging of line filter



Temperature rise due to varnish attached to the cooler


PE(Phosphate ester) Oil Application – EHC Oil

EHC oil is mainly hydrolyzed by water, and degradation process by oxidation, thermal deterioration, and the like. As a result of the degradation, varnishes and soots are produced, and the degree of formation of varnishes and soot can be managed as MPC (ΔL or $\Delta a + b$) values.


It is possible to quantify by measuring ΔL and $\Delta a + b$ of varnish and soot of phosphate ester control oil which can not be distinguished by MPC gravimetric method.

If ΔL and $\Delta a + b$ are high, the risk of varnish and soot is high. If $\Delta a + b$ is low and ΔL is high, it is contamination of soot.

When $\Delta a + b$ is low and ΔL is also low, the patch has a light yellow color, which means that the control oil is very good. Control oil oxidation is the main cause of oil deterioration and causes valve control failure and sticking. Carbide mainly causes valve wear, which causes leaking problems and valve sticking.



- Oxidation ($\Delta a + b \uparrow$): Oil with high ΔE and high $\Delta a + b$




✓ Bad valve control

✓ Valve Sticking

✓ Oil degradation

- Carbonization ($\Delta L \uparrow$): Oil with high ΔE and high ΔL



✓ Pressure issue for leakage

✓ Valve Sticking

Fully Complied to ASTM D7843-12



EHC Fluid improvement test method
is sourced from EPT. Cleanoil.com

ΔE (Normal MPC Test)

ΔL (Phosphate ester oils only)

Statistics

Notice

1. Alarm Limit

Normal A

Normal B

Caution

Warning

< 15

15 ~ 30

30 ~ 40

> 40

Save

Cancel

2. Calibration

X=94.81 Y=100.00 Z=107.30

Calibration

Data Transfer

3. Volume

ON

OFF

DIRECT_TEST_1 - ΔE

Measure

Back

Normal A

Normal B

Caut

<15

15 - 29

30 -

ΔE=0.46, ΔL=0.38, Δ

DIRECT_TEST_1 - ΔL

Measure

Back

ΔL

0.16

Δa+b

0.33

Δa+b

ΔL

6.3.10 Spectrophotometer, with capabilities of analyzing a standard 15mm target with a 0°/45° measuring a geometry, 10' observer, 10 um spectral intervals minimum resolution, the visible spectral range of 400-700 nm and CIELAB measuring indices.

D7843 - 12

6. Apparatus

6.1. Variance of apparatus, particularly with respect to filter measurement and vacuum pump setting, can affect the test results significantly.

6.2. When the user of this test method uses an alternate membrane filter, it is incumbent upon them to establish that the alternate filter will give equal results.

6.3. Required Apparatus:

6.3.1 Membrane Filter, 47 mm i.d. x 0.45 μm.

6.3.2 Forceps, smooth-tip.

6.3.3 Borosilicate Filter Holder.

6.3.4 Borosilicate Filtering Flask.

6.3.5 Wash Bottle equipped with 0.22 micron membrane.

6.3.6 Vacuum Source, capable of maintaining a vacuum of 71 kPa ± 5 kPa.

6.3.7 Graduated cylinder, 150-200 mL.

6.3.8 Beaker, 100-250 mL.

6.3.9 Spectrophotometer, with capabilities of analyzing a standard 15 mm target with a 0°/45° measuring geometry, 10' observer, 10 nm spectral intervals minimum resolution, the visible spectral range of 400-700 nm and CIELAB measuring indices.

6.3.10 Spectrophotometer, with capabilities of analyzing a standard 15mm target with a 0°/45° measuring a geometry, 10' observer, 10 um spectral intervals minimum resolution, the visible spectral range of 400-700 nm and CIELAB measuring indices.

7. Reagents and Materials

7.1. Petroleum Spirit (also known as petroleum ether or BP Petroleum Spirit 40/60) (Warning—Extremely Flammable. Hazardful if inhaled. Vapors are easily ignited by electrostatic discharges, causing flash fire), having boiling range from 35 to 60°C.

7.2. Coleman Lamp Fuel—Coleman Fuel is a complex mixture of light hydrocarbons (primarily aliphatic) produced by distillation of petroleum. Carbon number range is C3 to C5, and contains less than 0.001% benzene.

7.3. Purities of Reagents—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents conform to the specifications of the Committee of Analytical Reagents of the American Chemical Society when such specifications are available. Other grades may be used, provided it is first ascertained that the reagents are sufficiently high purity to permit use without lessening the accuracy of the determination.

8. Sampling, Test Specimens, and Test Unit

8.1. Using either Practice D6007 (manual sampling) or Practice D4177 (automatic sampling), obtain a representative sample of at least 50 mL of the material to be tested. The prepared sample container is an amber or other dark colored bottle to minimize potential UV reactions to the sample. Thickened or clear bottles can be used so long as the sample is protected from UV exposure.

8.2. The sample shall be heated to 60-65°C for 25-30 min, stored between 15-25°C, away from UV light for an incubation period of 6-16 hours. Samples that are washed prior to this aging period may produce fewer color bodies on the patch, thus, producing a lower ΔE value, and may lower the value of total analysis.

9. Procedure

9.1. Preparation of Sample and Materials:

9.1.1 Document the date and time at the beginning of the test.

9.1.2 Vigorously mix the sample for 15 ± minimum to resuspend inclusions uniformly. Visually inspect the inside of the bottle for evidence of material adherent to the surface of the bottle before sampling.

9.1.3 Thoroughly clean the bottle and use the bottle when the bottle is used.

9.1.4 Add approximately 30 mL ± 1 mL of petroleum ether to the bottle containing sample.

9.1.5 For sample for approximately 30 s to ensure that a complete solution (and complete mixing) is attained.

9.1.6 Pour sample into filter flask within 1-2 minutes of initial mixing.

9.2. Filtration Process:

9.2.1 Using forceps, mount the filter on the center of the filter holder.

9.2.2 Mount and securely clamp the filter flask to the filter.

9.2.3 Apply a vacuum and ensure a vacuum of less than 76 kPa is attained and held.

9.2.4 Remove the beaker twice with a maximum of 35 mL of petroleum ether and pour the slurry into the filter flask.

9.2.5 Repeat the slurry to completely flow through.

9.2.6 Carefully remove the clamp and filter. Wash any adhering inclusions from the flask onto the membrane with petroleum ether. Wash the membrane gently, particularly the edges, with petroleum ether from the wash bottle.

9.2.7 Carefully remove the vacuum.

9.2.8 Remove the clamp and filter holder.


9.2.9 Using forceps carefully remove the filter from the filter holder and place into a clean dry petri-dish. To facilitate handling, the membrane filter might be placed on clean glass rods in the petri-dish.

9.2.10 Dry the membrane by placing it in a low-level heat source free of ignition sources for flammable vapors, or in a dry (typically, 5 h) in a dust-free location. Dryness can be estimated by comparing the white color of the outer edge of the test membrane with a new membrane.

9.3. Color Determination of the Membrane Patch:

9.3.1 Transilluminate the instrument using a patch developed from clean solvent application of the method to establish the background color value.

9.3.2 Follow the standardization procedure defined by the instrument manufacturer.



WRTECTOR

Features	Specification
Appearance	
Size	214(W) x 306 (L) x 254 (H) / 5.5kg
Power	DC 220V with 24V, 5A
Measuring Principle	
Measuring Geometry	0°/45° measuring geometry (in full compliance with ASTM D7843)
Measurement Condition	Observer: CIE 10° Standard Observer
Light Source	LED Light
Receiver	Spectrum scan
Detector	Spectrophotometer
Measuring Time	3 Seconds
Operating Temperature	0° C ~ 50° C
Output Value	CIE delta E, delta L, a, b
Patch Color Capturing	YES
Interface	
OS	Linux
Analog Peripherals	7" Touch Screen LCD,

Additional Features:

- Sample Loading System
- Self-diagnosis and verification
- Output: Delta E and L can be output for MPC measurement for Delta E for Turbine and Delta L for EHC
- Unique calibration function
- Keyboard connection
- Automatic save and export of measured result values, USB storage
- Automatic diagnosis evaluation report function (option: printer)

